



For Battery Power Line II (BPL II) Li-ion battery (530.630)

STERRAD[®] and V-PRO[®] Systems Sterilization Guide

1. Preparation Prior to Reprocessing

- Before cleaning, remove all instruments and attachments from the power tool. Remove the battery casing from the handpiece and then remove the battery from the battery casing.
- Inspect battery for cracks or damage.

Note: Take care not to spray the contacts or touch both contacts at the same time with the damp cloth due to danger of short circuiting. Do not immerse batteries into water or clean them in a washer/disinfectant.

2. Cleaning and Disinfection

- A. To clean the battery, wipe it off with a clean, soft, and lint-free cloth dampened with deionized water and dry prior to processing.
- B. To disinfect the battery, wipe it off with a clean, soft and lint-free cloth dampened with a minimum of 70% alcohol-based disinfectant for thirty (30) seconds. A disinfectant that is VAH listed, EPA registered or locally recognized is recommended. This step has to be repeated two (2) additional times using a new, clean, soft and lint-free cloth dampened with a minimum 70% alcohol-based disinfectant each time. Follow the instructions provided by the manufacturer of the disinfectant.

3. Charging¹

- Place the battery into the Universal Battery Charger II² (05.001.204).
- Charge the battery. When the green LED illuminates, the battery is fully charged.
- Upon completion of charging the battery, wipe the battery with a minimum 70% alcohol-based disinfectant prior to wrapping and labeling.

4. Wrapping and Labeling

The batteries can either be wrapped or peel pouched. Follow the instructions for wrapping and labeling with the STERRAD Systems provided by Advanced Sterilization Products or with the V-PRO Systems provided by STERIS.

5. Sterilization

- Process the battery in the STERRAD or V-PRO Systems following the instructions provided by the legal manufacturer Advanced Sterilization Products or STERIS.
- The BPL II Li-Ion battery (530.630) has been validated with the following:

STERRAD and V-PRO Systems and Cycles

STERRAD[®] 50

STERRAD[®] 200 Short cycle

STERRAD[®] 100S Short cycle

STERRAD[®] NX[®] Standard cycle

STERRAD[®] 100NX[®] Standard cycle

V-PRO[®] 1 Lumen Cycle

V-PRO[®] 1 Plus Lumen Cycle, Non Lumen Cycle

V-PRO[®] maX Lumen Cycle, Non Lumen Cycle,
Flexible Cycle

V-PRO[®] 60 Lumen Cycle, Non Lumen Cycle

Note: No other sterilization methods are allowed.

6. Storage and Use³

- Battery can be used immediately or sterile stored in accordance with hospital's procedures.
- Rotate stock and recharge unused batteries, as necessary.
- Compared to nickel based batteries the self-discharge rate of the BPL II Li-Ion battery is less than half during storage⁴.

Warning:

- **Never expose batteries to temperature over 60°C/140°F. The maximum exposure time at 60°C/140°F is 72 hours.**
- **Do not dismantle, open, short-circuit or shred batteries.**
- **Do not leave the battery empty of charge as this will reduce its lifetime or damage the battery.**

¹ Universal Battery Charger II with firmware version 14.0 or higher is required to charge BPL II Li-ion batteries.

² For detailed information on the Universal Battery Charger II, please refer to the Instructions for Use for the Universal Battery Charger II (DSEM/PWT/1114/0050).

³ For detailed information regarding Storage and Use, please refer to the Instructions for Use for the Battery Power Line II (DSEM/PWT/1114/0051).

⁴ Data on file at DePuy Synthes.

7. STERRAD or V-PRO Sterilized Battery Insertion

- The scrubbed person performs all of the steps according to Figure 1–8.
- Open the lid of the battery casing (530.690) by simultaneously pressing both battery casing locks (Fig. 1). Ensure the lid of the battery casing is fully open (Fig. 2) and facing towards the scrubbed person (Fig. 3).
- Insert the sterile battery into the battery casing (Fig. 4) and press it completely into the battery casing to ensure a correct seat (Fig. 5).
- Close the battery casing by simultaneously pressing both battery casing locks and closing the lid (Fig. 6 and Fig. 7).
- Attach the battery pack into the Battery Power Line II handpiece (Fig. 8).

After usage, remove the battery casing from the handpiece and open the lid. The circulating person then removes the battery. Ensure that the battery does not touch the exterior of the battery casing to avoid contaminating the battery. Put the battery immediately in the charger.

Notes:

- Ensure that the STERRAD or V-PRO sterilized battery does not come into contact with any liquids or other substances whilst being inserted.
- For detailed information please refer to the Instructions for Use for the Battery Power Line II (DSEM/PWT/1114/0051).
- Following the introduction of the STERRAD or V-PRO sterilization methods, the markings on the BPL II Li-Ion batteries (530.630) have been changed, including replacing “DO NOT STERILIZE” with “DO NOT AUTOCLAVE MAX. 60°C/140°F”. However, all BPL II Li-Ion batteries (530.630) can be STERRAD or V-PRO sterilized, regardless of their markings.



Figure 1



Figure 2



Figure 3



Figure 4



Figure 5



Figure 6



Figure 7



Figure 8

STERRAD® is a registered trademark of Advanced Sterilization Products (ASP).

V-PRO® is a registered trademark of STERIS Corporation.

This publication is not intended for distribution in the USA.

For full user instructions and precautions please consult the product IFU. Please contact your DePuy Synthes sales representative for more information.

All surgical techniques are available as PDF files at www.depuysynthes.com/ifu



Synthes GmbH
Eimattstrasse 3
4436 Oberdorf
Switzerland
Tel: +41 61 965 61 11
Fax: +41 61 965 66 00
www.depuysynthes.com