
Instructions for Use Instruments with a Measuring Function

This instruction for use is not intended for distribution in the USA.

Instructions for Use

Instruments with a Measuring Function

Important note for medical professionals and OR staff: These Instructions for Use do not include all of the information necessary for selection and use of a device. Please read the instructions for use, the DePuy Synthes brochure "Important Information" carefully before use. Ensure that you are familiar with the appropriate surgical technique.

Materials

Aluminum
PEEK
PPSU
Silicone Rubber
Stainless Steel
Ti Al6 V4

Intended use

Instruments with a Measuring Function are intended to measure quantitatively an anatomical parameter of the human body.

Intended User

This device is intended to be used by qualified health care professionals e.g. surgeons, physicians, radiologists, operating room staff, and professionals involved in preparation of the device. All personnel handling the device should be fully aware of the intended use of the products, the applicable surgical techniques and the DePuy Synthes "Important Information" leaflet.

Potential Adverse Events

DePuy Synthes manufactures surgical measuring instruments intended to prepare the site and aide in implantation of DePuy Synthes implants. The adverse events/side effects are based upon the implant devices rather than the instruments. Specific adverse events/side effects for the implants can be found in the respective implant Instructions For Use.

Measuring Device

Measuring Devices are intended to measure quantitatively an anatomical parameter of the human body.

Part	Measuring range (mm)	Degree of Accuracy (mm)
Calipers		
324.060	0–180	±0.2
389.186	0–180	±0.2
03.501.065	0–60	±0.5
03.501.074	0–40	±0.5
Depth Gauges		
319.003	0–25	±0.85
319.004	0–30	±1.25
319.005	0–43	±0.85
319.006	0–50	±1.25
319.010	0–60	±0.7
319.011	0–110	±0.65
319.060	0–40	±0.6
319.090	4–110	±0.5
319.091	10–150	±1.1
319.100	14–110	±0.7
319.110	0–24	±0.5
319.520	0–44	±0.4
319.530	0–44	±1
323.040	18–114	±0.8
323.041	Only used in conjunction with 323.040	
355.790	14–90	±0.45
356.835	18–110	±1.1
357.402	12–110	±0.8
357.789	20–78	±0.85
357.790	14–100	±0.5

Part	Measuring range (mm)	Degree of Accuracy (mm)
357.790–EXS	14–100	±0.5
357.791	14–100	±0.45
387.292	0–50	±0.75
03.010.019	18–110	±1.1
03.010.072	18–110	±1.1
03.010.428	18–110	±1.1
03.010.494	10–100	±0.6
03.019.017	20–98	±0.5
03.019.029	20–98	±0.5
03.025.052	20–110	±1.1
03.108.026	20–110	±0.5
03.113.028	10–100	±0.9
03.118.007	0–100	±0.9
03.120.049	10–60	±0.6
03.122.052	24–74	±0.15
03.130.250	0–43	±0.6
03.161.028	0–50	±0.75
03.168.017	0–100	±1.1
03.305.005	40–90	±0.95
03.501.001	0–50	±1
03.503.036	0–44	±1
03.503.085	0–44	±0.5

Direct Measuring Devices

311.690	40–150	±0.8
311.690–EXS	40–150	±0.8
311.720	70–120	±0.6
319.150	8–70	±0.75
319.155	20–80	±0.75
319.170	20–80	±0.75
319.210	10–150	±0.95
319.700	30–180	±0.85
319.701	25–140	±1.1
319.702	8–50	±0.95
319.703	0–40	±0.9
323.029	6–32	±0.4
323.034	6–32	±0.5
323.060	10–60	±1.05
323.061	6–58	±0.5
324.037	18–85	±4.1
324.208	0–140	±1.3
338.050	40–170	±1.15
338.170	40–150	±1
338.329	75–150	±0.9
351.717	160–480	±3.7
356.829	70–130	±1.4
357.042	60–125	±0.85
357.385	80–130	±1.9
357.430	70–150	±0.4
358.698	34–54	±1
360.255	160–340	±3.4
387.550	28–50	±1.5
03.631.521	25–100	±1.483
03.010.083	40–100	±1.55
03.010.085	40–130	±0.85
03.010.090	34–54	±0.6
03.010.106	18–100	±1
03.010.429	18–100	±0.8

Part	Measuring range (mm)	Degree of Accuracy (mm)
03.010.492	40–100	± 1.55
03.010.493	40–130	± 0.85
03.037.020	70–130	± 0.9
03.037.027	70–130	± 0.4
03.037.036	240–480	± 3.7
03.108.003	10–80	± 2.1
03.108.037	10–50	± 1.3
03.110.000	10–30	± 0.9
03.110.006	0–40	± 1
03.111.000	0–30	± 0.3
03.111.005	0–40	± 0.6
03.168.003	50–140	± 0.8
03.207.004	35–184	± 0.8
03.226.002	5–40	± 0.95
03.226.008	10–40	± 1.1
03.226.030	18–110	± 0.75
03.227.030	30–160	± 0.75
03.231.017	25–145	± 1.6
Osteotomy Correction Devices		
395.000	0–70/4°–20°	± 0.41/4°
395.001	5–16	± 0.9
03.108.008	0–100	± 0.1
03.108.039	0–100	± 0.1
03.211.009	1.5–7	± 0.3
Reduction Instruments		
313.354	0–70	± 1.2
399.003	0–30	± 0.5
Rulers		
333.370	0–180	± 0.1
03.401.083	0–250	± 0.1
Height Measuring Devices		
324.092	0–120	± 1.5
03.661.010	20–150	± 0.45

Precautions

Measuring devices may be used to measure various anatomical parameters in the human body and in the measurement of various implantable medical devices. Device specific precautions and warnings related to measuring devices are described in the corresponding Instructions for Use for those specific implantable devices.

Health care professionals should be qualified by appropriate training in the use of measuring devices. Measurement error can potentially result in implant failure from selecting the wrong size of implant.

Measured parameters are only valid where measuring devices are processed and maintained per DePuy Synthes "Important Information".

Take care to avoid damaging the nerve and vessel bundles, soft tissue and organs present in the human body where measuring devices are being used to measure an anatomical parameter.

These devices can break during use (when subjected to excessive forces or outside the recommended surgical technique). While the surgeon must make the final decision on removal of the broken part based on associated risk in doing so, we recommend that whenever possible and practical for the individual patient, the broken part should be removed.

Combination of medical devices

DePuy Synthes has not tested compatibility with devices provided by other manufacturers and assumes no liability in such instances.

Treatment before device is used

DePuy Synthes products supplied in a non-sterile condition must be cleaned and steam-sterilized prior to surgical use. Prior to cleaning, remove all original packaging. Prior to steam-sterilization, place the product in an approved wrap or container. Follow the cleaning and sterilization instruction given by the DePuy Synthes brochure "Important Information".



Non-Sterile

Clinical Processing of the Device

Detailed instructions for reprocessing of reusable devices, instrument trays and cases are described in the DePuy Synthes brochure "Important Information". Assembly and disassembly instructions of instruments "Dismantling multipart instruments" can be downloaded from <http://emea.depuyssynthes.com/hcp/reprocessing-care-maintenance>.

Disposal

Devices must be disposed of as a healthcare medical device in accordance with hospital procedures.

Not all products are currently available in all markets.



Synthes GmbH
Eimattstrasse 3
4436 Oberdorf
Switzerland
Tel: +41 61 965 61 11
Fax: +41 61 965 66 00
www.depuyssynthes.com