
Instructions for Use Plate and Screw Implants

This instruction for use is not intended for
distribution in the USA.

Instructions for Use

Plate and Screw Implants

Associated device systems with these instructions for use:

2.4 mm Cannulated Screw
2.4 mm Variable Angle LCP Volar Extra-Articular Distal Radius System
2.4/2.7 mm Locking Tarsal Plates
Angled Blade Plates for Adults
Angular Stable X-Plate and 2-Hole Plate
Cannulated Pediatric Osteotomy System (CAPOS)
Cannulated Screws 3.0/3.5/4.0/4.5/6.5/7.0/7.3
DCP and LC-DCP Systems
DHS/DCS System
Femoral Neck System
HCS 1.5
HCS 2.4/3.0
HCS 4.5/6.5
LCP Anterolateral Distal Tibia Plate 3.5
LCP Clavicle Hook Plate
LCP Compact Foot / Compact Hand
LCP Compact Hand
LCP Compact Hand 1.5
LCP Condylar Plate 4.5/5.0
LCP DHHS
LCP Dia-Meta Volar Distal Radius Plates
LCP Distal Fibula Plates
LCP Distal Humerus Plates
LCP Distal Radius System 2.4
LCP Distal Tibia Plate
LCP Distal Ulna Plate
LCP Extra-articular Distal Humerus Plate
LCP Hook Plate 3.5
LCP Locking Compression Plate
LCP Low Bend Medial Distal Tibia Plates 3.5 mm
LCP Medial Proximal Tibial Plate 3.5
LCP Medial Proximal Tibial Plate 4.5/5.0
LCP Metaphyseal Plate for distal medial tibia
LCP Metaphyseal Plates
LCP Olecranon Plate
LCP Pediatric Condylar Plate 90°, 3.5 and 5.0
LCP Pediatric Plate System
LCP Periarticular Proximal Humerus Plate 3.5
LCP Posterior Medial Proximal Tibial Plate 3.5
LCP Proximal Femoral Hook Plate 4.5/5.0
LCP Proximal Femoral Plate 4.5/5.0
LCP Proximal Radius Plates 2.4
LCP Proximal Tibial Plate 3.5
LCP Proximal Tibial Plate 4.5/5.0
LCP Superior Anterior Clavicle Plate
LCP Superior Clavicle Plate
LCP Ulna Osteotomy System 2.7
LCP Volar Column Distal Radius Plates 2.4 mm
LCP Wrist Fusion Set
LISS and LCP DF
LISS and LCP PLT
Locking Attachment Plate
Midfoot Fusion Bolt Ø 6.5 mm
Orthopedic Foot Instruments
Pelvic Implants and Instruments
Periarticular Aiming Arm Instruments for LCP Proximal Tibial Plate 4.5/5.0
PHILOS and PHILOS Long
PHILOS WITH AUGMENTATION
Quadrilateral Surface Plates 3.5
Rotation Correction Plates 1.5 and 2.0
Sacral Bars
Slipped Capital Femoral Epiphysis (SCFE) Screw System
Spring Plates 3.5

Standard DHS Lag Screw with LCP DHHS Sideplate
The Calcaneal Plate
The Locking Calcaneal Plate
TomoFix
TomoFix Medial Distal Femur (MDF)
TomoFix Medial High Tibial Plate (MHT)
VA-LCP® MEDIAL COLUMN FUSION PLATES 3.5
VA LOCKING CALCANEAL PLATES 2.7
VA-LCP Ankle Trauma System 2.7/3.5
VA-LCP Anterior Clavicle Plate
VA-LCP Condylar Plate 4.5/5.0
VA-LCP Distal Humerus Plates 2.7/3.5
VA-LCP Olecranon Plates 2.7/3.5
VA-LCP Proximal Tibial Plate 3.5
VA-Locking Intercarpal Fusion System
Variable Angle LCP 1st MTP Fusion Plates 2.4/2.7
Variable Angle LCP Dorsal Distal Radius Plate 2.4
Variable Angle LCP Forefoot/Midfoot System 2.4/2.7
Variable Angle LCP Mesh Plate 2.4/2.7
Variable Angle LCP Opening Wedge Plates 2.4/2.7
Variable Angle LCP Tarsal Plates 2.4/2.7
Variable Angle LCP TMT Fusion Plates 2.4/2.7
Variable Angle LCP Two-Column Volar Distal Radius Plate 2.4
Variable Angle LCP Two-Column Volar Distal Radius Plate 2.4/2.7, Extra-long
Variable Angle LCP Volar Rim Distal Radius Plate 2.4
Variable Angle Locking Hand System

Please read these instructions for use, the Synthes brochure "Important Information" carefully before use. Ensure that you are familiar with the appropriate surgical technique.

Plate and Screw Implants consist of various plates and screws to be implanted which are single packed, and available sterile and/or non-sterile.

Important note for medical professionals and OR staff: These instructions for use do not include all of the information necessary for selection and use of a device. Please see full labeling for all necessary information (corresponding Surgical Technique Guide, Important Information and device-specific label).

Material(s)

Material(s):	Standard(s):
Stainless Steel – 316L	ISO 5832-1
Stainless steel – 22-13-5	ASTM F 1314
TiCP	ISO 5832-2
CoCrMo alloy	ISO 5832-12
Titanium alloy:	
Ti-6Al-7Nb (TAN)	ISO 5832-11
Ti-6Al-4V (TAV)	ISO 5832-3
Ti-15Mo	F 2066

Intended use

Plate and Screw Implants are intended for temporary fixation, correction or stabilization of bones in various anatomical regions.

Indications

Please refer to the table at the end of this IFU.

Contraindications

Please refer to the table at the end of this IFU.

Potential risks

As with all major surgical procedures, risks, side effects and adverse events can occur. While many possible reactions may occur, some of the most common include: Problems resulting from anesthesia and patient positioning (e.g. nausea, vomiting, dental injuries, neurological impairments, etc.), thrombosis, embolism, infection, excessive bleeding, iatrogenic neural and vascular injury, damage to soft tissues incl. swelling, abnormal scar formation, functional impairment of the musculoskeletal system, Sudeck's disease, allergy/hypersensitivity reactions, and side effects associated with hardware prominence, malunion, non-union.

Sterile device

STERILE R Sterilized using irradiation

Store implants in their original protective packaging, and do not remove them from the packaging until immediately before use. Do not use when packaging is damaged.

Prior to use, check the product expiration date and verify the integrity of the sterile packaging. Do not use, if the package is damaged or date of expiration has passed.



Do not re-sterilize

Implantable devices labeled with "Do not re-sterilize" symbol must not be re-sterilized because re-sterilization may compromise the structural integrity of the device and/or may lead to device failure. Re-sterilization of implantable devices can result in product not being sterile, and/or not meeting performance specifications and/or altered material properties.

Single-use device



Do not re-use

Indicates a medical device that is intended for one use, or for use on a single patient during a single procedure.

Re-use or clinical reprocessing (e.g. cleaning and re-sterilization) may compromise the structural integrity of the device and/or lead to device failure which may result in patient injury, illness or death.

Furthermore, reuse or reprocessing of single-use devices may create a risk of contamination e.g. due to the transmission of infectious material from one patient to another. This could result in injury or death of the patient or user.

Contaminated implants must not be reprocessed. Any Synthes implant that has been contaminated by blood, tissue, and/or bodily fluids/matter should never be used again and should be handled according to hospital protocol. Even though they may appear undamaged, the implants may have small defects and internal stress patterns that may cause material fatigue.

Precautions

For general precautions consult "Important Information".

For application specific precautions related to Plate and Screw Implants it is mandatory to consult the corresponding Surgical Technique Guide (www.depuysynthes.com/ifu) of the product system being used.

Warnings

For general warnings consult "Important Information".

For application specific warnings related to Plate and Screw Implants it is mandatory to consult the corresponding Surgical Technique Guide (www.depuysynthes.com/ifu) of the product system being used.

Combination of medical devices

Synthes has not tested compatibility with devices provided by other manufacturers and assumes no liability in such instances.

Magnetic Resonance environment

When a device has been evaluated for use in the MR environment, MRI information will be found in the surgical technique at www.depuysynthes.com/ifu.

Treatment before device is used

Synthes products supplied in a non-sterile condition must be cleaned and steam-sterilized prior to surgical use. Prior to cleaning, remove all original packaging. Prior to steam-sterilization, place the product in an approved wrap or container. Follow the cleaning and sterilization instruction given by the Synthes "Important Information".

Clinical Processing/reprocessing of the device

Detailed instructions for processing of implants and reprocessing of reusable devices, instrument trays and cases are described in the Synthes brochure "Important Information". Assembly and disassembly instructions of instruments "Dismantling multi-part instruments" can be downloaded from <http://emea.depuysynthes.com/hcp/reprocessing-care-maintenance>

Systems	Indications	Contraindications
2.4/2.7 mm Locking Tarsal Plates	The Synthes 2.4 mm/2.7 mm locking tarsal plates are intended for the fixation of fractures, osteotomies, nonunions, replantations, and fusions of the Cuboid, Tarsal and Navicular bones, particularly in osteopenic bone.	No contraindication specific to these devices.
2.4 mm Cannulated Screw	<ul style="list-style-type: none">– Intra-articular fractures of the carpals, metacarpals, tarsals and metatarsals– Fixation of small bone fragments– Bunionectomies and osteotomies– Arthrodeses of small joints	No contraindication specific to these devices.
2.4 mm Variable Angle LCP Volar Extra-Articular Distal Radius System	The 2.4 mm Variable Angle LCP Volar Extra-Articular Distal Radius Plates are indicated for fixation of intra and extra-articular fractures and osteotomies of the distal radius.	No contraindication specific to these devices.
Angled Blade Plate for Adults	<p>130° Angled Blade Plate Fractures and revisions of the proximal third of the femur in skeletally mature patients.</p> <p>Condylar Plates, 95° Fractures and revisions of the proximal and distal third of the femur in skeletally mature patients.</p> <p>Osteotomy Plates 90°/100°/110°/120°/130° Hip Plate Osteotomies on the proximal femur in skeletally mature patients.</p>	No contraindication specific to these devices.

Systems	Indications	Contraindications
Angular Stable X-Plate and 2-Hole Plate	<p>X-plate</p> <p>Arthrodesis</p> <ul style="list-style-type: none"> – MTP – MTC – Talonavicular – Calcaneo-Cuboid <p>Proximal osteotomies</p> <ul style="list-style-type: none"> – Crescentic – Ludloff – Mau – Open wedge – Closing wedge <p>Calcaneal osteotomies</p> <p>Fractures</p> <ul style="list-style-type: none"> – Single – Multifragmental <p>2-hole plate</p> <ul style="list-style-type: none"> – Akin osteotomy 	No contraindication specific to these devices.
Cannulated Pediatric Osteotomy System (CAPOS)	<p>The CAPOS System is indicated for use in infants, toddlers, children, adolescents, and small-stature adult patients.</p> <p>Specific indications include:</p> <ul style="list-style-type: none"> – Intertrochanteric derotation and varus osteotomies <p>Osteotomy Plates</p> <ul style="list-style-type: none"> – Mini Infant Hip Plate, Infant Hip Plate 3.5, 90°, Child Hip Plate 3.5, 90°, Child Hip Plate 4.5, Hip Plate 80°, 90° and 100° (Adolescent) – Intertrochanteric valgus osteotomies Plates – 95° Condylar Plates (Adolescent and Small Stature Adult) – Femoral neck and pertrochanteric fractures – Plates – 130° Angled Blade Plates (Adolescent and Small Stature Adult) 	No contraindication specific to these devices.
Cannulated Screws 3.0/3.5/4.0/4.5/6.5/7.0/7.3	<p>CSS 3.0</p> <p>Fixation of fractures of the forearm, hand and foot, e.g.:</p> <ul style="list-style-type: none"> – fractures and arthrodeses of the carpals and metacarpals – fractures of the distal radius and radial head – metatarsal fractures <p>CSS 3.5</p> <p>Fixation of fractures with small fragments, e.g.:</p> <ul style="list-style-type: none"> – wrist fractures – metacarpal and metatarsal fractures and fixation in metacarpal and metatarsal osteotomies – tarsal fractures – transcondylar humeral fractures in children <p>CSS 4.0</p> <p>Fixation of fractures with medium fragments, e.g.:</p> <ul style="list-style-type: none"> – tarsal and metatarsal fractures and fixation in metatarsal and phalangeal osteotomies – tarsometatarsal and metatarsophalangeal arthrodesis – ligament fixations – hallux valgus corrections <p>CSS 4.5</p> <p>Fixation of fractures with medium fragments, e.g.:</p> <ul style="list-style-type: none"> – malleolar fractures – pilon tibial fractures – fractures of the calcaneus and talus – tibial plateau fractures – carpal and tarsal arthrodesis <p>CSS 6.5, 7.0 and 7.3</p> <p>Fixation of fractures with large fragments, e.g.:</p> <ul style="list-style-type: none"> – femoral neck fractures – intercondylar femoral fractures – epiphyseolysis of the femoral head – ankle arthrodeses – iliosacral dislocations 	No contraindication specific to these devices.

Systems	Indications	Contraindications
DCP and LC-DCP Systems	<p>This surgical technique applies to the below Synthes DCP and LC-DCP systems and plate lines:</p> <p>Mini Fragment System 1.5, 2.0, 2.7</p> <ul style="list-style-type: none"> – Fractures of the middle and distal phalanges and tarsals – Fractures of the metacarpals and metatarsals – Osteotomies and arthrodeses on the hand and foot – Fractures of the distal radius (double-plate technique) <p>Compact Hand 1.0, 1.3, 1.5, 2.0, 2.4</p> <p>Implant sizes 1.0/1.3:</p> <ul style="list-style-type: none"> – Fixation of small fragments – Fractures of the distal and intermediate phalanges – Avulsion fractures <p>Implant sizes 1.5:</p> <ul style="list-style-type: none"> – Fractures of the phalanges and metacarpals – Capturing of fragments with lag screw technique <p>Implant sizes 2.0/2.4:</p> <ul style="list-style-type: none"> – Fractures of the phalanges and metacarpals – Capturing of fragments with lag screw technique <p>Cloverleaf Plates 3.5</p> <ul style="list-style-type: none"> – Distal tibia for comminuted fractures to buttress its medialside – Proximal humerus for comminuted fractures of the humeralhead <p>One-third tubular plate 3.5</p> <ul style="list-style-type: none"> – Fractures of smaller sized bones such as fibula, humerus and ulna. <p>LC-DCP Plate 3.5, DCP Plate 3.5, T-Plate 3.5</p> <ul style="list-style-type: none"> – Fracture fixation and fixation after for example osteotomies, malunions, nonunions including but not limited to distal radius, proximal and distal tibia, proximal humerus, and clavicle. <p>LC-DCP Plate 4.5, DCP Plate 4.5, T-Plate 4.5, T-Buttress Plate 4.5, L-Buttress Plate 4.5</p> <ul style="list-style-type: none"> – Fractures and osteotomies of large bones such as femur, tibia and humerus <p>Semi-Tubular Plate 4.5</p> <ul style="list-style-type: none"> – Fractures and osteotomies of smaller sized bones such as humerus, radius, ulna, clavicle, fibula, tibia and pelvis <p>Condylar Buttress Plate 4.5</p> <ul style="list-style-type: none"> – Buttressing of multifragmentary distal femur fractures – Supracondylar fractures – Intra-articular and extra-articular condylar fractures – Malunions and nonunions of the distal femur – Periprosthetic fractures <p>Lateral Tibial Head Buttress Plate 4.5</p> <ul style="list-style-type: none"> – Indicated for the stabilization of fractures of the proximal tibia. These include proximal shaft fractures, metaphyseal fractures, intra-articular fractures, periprosthetic fractures. <p>Proximal Tibial Plate 4.5</p> <p>Indicated for nonunions, malunions and fractures of the proximal tibia including simple, comminuted, lateral wedge, depression, medial wedge, bicondylar and combinations of lateral wedge and depression fractures.</p>	<p>No contraindication specific to these devices.</p>
DHS/DCS System	<p>The DHS system including all combinations of DHS Screw, DHS Blade, DHS plate with DCP holes, LCP DHS plate and LCP DHS with collar</p> <p>Indications DHS</p> <ul style="list-style-type: none"> – Ptertrochanteric fractures of type 31-A1 and 31-A2 – Intertrochanteric fractures of type 31-A3 – Basilar neck fractures 31-B (DHS Screw in conjunction with an antirotation screw) – Subtrochanteric fractures <p>Trochanter Stabilizing Plate Indications LTSP/ULTSP/TSP</p> <ul style="list-style-type: none"> – Unstable ptertrochanteric fractures of type 31-A2 and 31-A3, especially multifragmentary fractures with separated or longitudinally split greater trochanter <p>Indications DCS</p> <ul style="list-style-type: none"> – Proximal femur: Very proximally located, purely subtrochanteric fractures of types 32-A and 32-B – Distal femur: Fractures of type 33-A (extra-articular, supracondylar) and fractures of type 33-C (fully articular fractures) 	<p>Contraindications DHS</p> <p>The DHS is not to be used in cases where there is a high incidence of:</p> <ul style="list-style-type: none"> – Sepsis – Malignant primary or metastatic tumors – Material sensitivity – Compromised vascularity <p>Contraindications DCS</p> <ul style="list-style-type: none"> – Ptertrochanteric fractures or trochanteric fractures with subtrochanteric expansion (31-A3)

Systems	Indications	Contraindications
Sterile Distal Radius Kit	Variable Angle LCP® Two-Column Volar Distal Radius Plates 2.4 are indicated for the fixation of intra- and extra-articular fractures and osteotomies of the distal radius.	No contraindication specific to these devices.
Femoral Neck System (FNS)	Femoral neck fractures (AO type 31-B)	<ul style="list-style-type: none"> – Pertrochanteric fractures (AO type 31-A1 and 31-A2) – Intertrochanteric fractures (AO type 31-A3) – Subtrochanteric fractures <p>Additionally, this system should not be used for cases where there is a high incidence of:</p> <ul style="list-style-type: none"> – Sepsis – Malignant primary or metastatic tumors – Material sensitivity – Compromised vascularity
HCS 1.5	<ul style="list-style-type: none"> – Fixation of intra- and extra-articular fractures and non-unions of small bones and small bone fragments – Arthrodesis of small joints – Osteochondral fractures – Osteotomies – Avulsion fractures 	No contraindication specific to these devices.
HCS 2.4/3.0	<ul style="list-style-type: none"> – Fixation of intra-articular and extra-articular fractures and non-unions of small bones and small bone fragments – Arthrodeses of small joints – Bunionectomies and osteotomies <p>Examples include, but are not limited to scaphoid and other carpal bones, metacarpals, tarsals, metatarsals, patella, ulnar styloid, capitellum, radial head and radial styloid.</p>	No contraindication specific to these devices.
HCS 4.5/6.5	<p>HCS 4.5 Fracture, osteoarthritis, or deformity of small to large bones. Examples:</p> <ul style="list-style-type: none"> – Calcaneus – Talus – Metatarsus – Distal and proximal tibia – Distal femur – Proximal humerus <p>HCS 6.5 Fracture, osteoarthritis, or deformity of small to large bones. Examples:</p> <ul style="list-style-type: none"> – Calcaneus – Talus – Distal and proximal tibia – Distal femur 	No contraindication specific to these devices.
LCP Anterolateral Distal Tibia Plate 3.5	The LCP Anterolateral Distal Tibia Plate 3.5 is indicated for: <ul style="list-style-type: none"> – Extra-articular and simple intra-articular distal tibia fractures – Distal tibia fracture, percutaneous or reducible by limited arthro-tomy – Distal tibia fracture extending into the diaphyseal area 	No contraindication specific to these devices.
LCP Clavicle Hook Plate	<ul style="list-style-type: none"> – Lateral clavicle fractures: Neer type II or Jäger and Breitner type II – Acromioclavicular joint dislocation Type: Tossy III or Rockwood III to V 	<ul style="list-style-type: none"> – Stable lateral clavicle fractures – Tossy Type I and II – Rockwood Type I and II – Acute infection
LCP Compact Foot/ Compact Hand	<p>Indications for implants of size 2.0 and 2.4 include:</p> <ul style="list-style-type: none"> – Fractures of the phalanges – Fractures of the metacarpals and metatarsals (II–V) – Fractures of the distal radius (double-plate technique) – Osteotomies and arthrodeses on the hand and foot(e.g. TMT [II–V] fusions) – Subcapital radial head fracture – As an additional implant with small fragments <p>Indications for implants of size 2.7 include:</p> <ul style="list-style-type: none"> – Fractures of metatarsal I – Fractures of the tarsals – MTP 1 fusions – Osteotomies and arthrodeses of the tarsals(e.g. calcaneo-cuboidal fusion) 	No contraindication specific to these devices.

Systems	Indications	Contraindications
LCP Compact Hand	<p>Compact Hand 1.0/1.3 General: – Fixation of small fragments – Avulsion fractures and fractures of the distal and middle phalanges</p> <p>Straight plate: – Shaft fractures of phalanges</p> <p>Strut plate: – Comminuted fractures and replantations of middle or proximal phalanges T-plate and</p> <p>Y-plate: – Condylar fractures of phalanges</p> <p>Single cortex screws: – Capturing of fragments with lag screw technique</p> <p>LCP Compact Hand 1.5 – Fracture fixation of the phalanges and metacarpals – Osteotomies – Arthrodeses – Replantations and reconstructions of phalanges and metacarpals, particularly in osteopenic bone</p> <p>LCP Compact Hand 2.0 – Fractures of the phalanges, metacarpals and wrist bones – Osteotomies and arthrodeses of the interphalangeal joints – Fractures of distal radius (double-plate technique) – Subcapital radial head fractures</p> <p>LCP Compact Hand 2.4 – Fractures of the metacarpal and wrist bones – Fractures of distal radius (double-plate technique) – Osteotomies and arthrodeses on the hand – Subcapital radial head fractures</p>	No contraindication specific to these devices.
LCP Compact Hand 1.5	– Fracture fixation of the phalanges and metacarpals – Osteotomies – Arthrodeses – Replantations and reconstructions of phalanges and metacarpals, particularly in osteopenic bone	No contraindication specific to these devices.
LCP Condylar Plate 4.5/5.0	– Buttressing of multifragmentary distal femur fractures – Supracondylar fractures – Intra-articular and extra-articular condylar fractures – Malunions and nonunions of the distal femur – Periprosthetic fractures – Fractures in normal or osteopenic bone	No contraindication specific to these devices.
LCP DHHS	The LCP DHHS is indicated for the following fractures of the proximal femur: – Intertrochanteric fractures – Basilar neck fractures – Pertrochanteric fractures LCP DHHS is indicated for stable and unstable fractures in which a stable medial buttress can be reconstructed.	No contraindication specific to these devices.
LCP Dia-Meta Volar Distal Radius Plates	LCP Dia-Meta Volar Distal Radius Plates are indicated for fractures, osteotomies, and nonunions of the radius.	No contraindication specific to these devices.
LCP Distal Fibula Plates	The LCP Distal Fibula Plates are intended for fixation of fractures, osteotomies and non-unions of the metaphyseal and diaphyseal region of the distal fibula, especially in osteopenic bone.	No contraindication specific to these devices.
LCP Distal Humerus Plates	Indications for Distal Humerus Plate – Intraarticular fractures of the distal humerus – Supracondylar fractures of the distal humerus – Non-unions of the distal humerus	(for DHP and for Metaphyseal plate) – Acute infections – Children in the growth phase
LCP Distal Radius System 2.4	Displaced extra-articular and intra-articular distal radius fractures and corrective osteotomies of the distal radius. Dorsal approach – Dorsally displaced fractures – Extra-articular fractures with metaphyseal defect (AO classification 23-A3) – Open joint reconstruction (AO classification 23-C1, C2, C3) – Combination of distal radius with carpal and metacarpal fractures – Corrective osteotomies Palmar approach– Reversed Barton – Palmarly displaced extra-articular fractures (Goyrand- Smith) – Dorsally displaced extra-articular (Colles) and articular fractures – Extra-articular fractures with extension into the shaft (extra-long plates)	No contraindication specific to these devices.

Systems	Indications	Contraindications
LCP Distal Tibia Plate	<ul style="list-style-type: none"> – Extra-articular and simple intra-articular distal tibial fractures – Distal tibial fractures, percutaneous or reducible by limited arthro-tomy – Distal tibial fracture extending into the diaphyseal area 	No contraindication specific to these devices.
LCP Distal Ulna Plate	<p>Fractures of the distal ulna which result in instability of the distal radioulnar joint.</p> <p>Fractures of the ulna where the articular surface is either displaced, rotated or tilted.</p> <p>Comminuted extra-articular fractures of the ulnar neck threatening stable congruency of the distal radioulnar joint.</p>	No contraindication specific to these devices.
LCP Extra-articular Distal Humerus Plate	<ul style="list-style-type: none"> – Extra-articular fractures of the distal humerus – Malunions of the distal humerus – Non-unions of the distal humerus 	No contraindication specific to these devices.
LCP Hook Plate 3.5	<ul style="list-style-type: none"> – Simple fractures of the olecranon (AO Types 21–B1, 21–B3, 21–C1) – Osteotomies of the olecranon for distal humerus fracture treatment – Avulsion fractures of the distal tibia and fibula 	No contraindication specific to these devices.
LCP Locking Compression Plate	<p>The indications for each system will be presented individually.</p> <p>Small Fragment plates (3.5) LCP Plates 3.5 Fixation of small bone fragments using the five standard AO plating principles (buttress plate, neutralization plate, tension band plate, bridge plate, compression plate).</p> <p>LCP Reconstruction Plates 3.5 Fixation of small bone fragments using the five standard AO plating principles (buttress plate, neutralization plate, tension band plate, bridge plate, compression plate).</p> <p>LCP T-Plates 3.5 Fixation of small bone fragments using the five standard AO plating principles (buttress plate, neutralization plate, tension band plate, bridge plate, compression plate). Fracture fixation and fixation after osteotomies, malunions, non-unions; e.g. including but not limited to distal radius, proximal and distal tibia, proximal humerus, clavicular.</p> <p>LCP One-third Tubular Plates 3.5 The LCP One-third Tubular Plates 3.5 are indicated for plating of fractures of long and small bones. The plate should only be used for load-sharing purposes, e.g. buttressing, tension banding, neutralization or compression.</p> <p>LCP Methaphyseal 3.5 The LCP Metaphyseal Plates are indicated to fix extra-articular fractures in the metaphyseal area that can extend into the shaft area. The 3.5 mm plates are indicated to fix fractures of the distal humerus and distal fibula.</p> <p>Large Fragment plates (4.5/5.0) LCP Methaphyseal 3.5/4.5/5.0 The LCP Metaphyseal Plates are indicated for Extra-articular fractures of the metaphyseal area that extend into the shaft area.</p> <p>The 4.5/5.0 plates are indicated to fix fractures of the proximal humerus and distal tibial.</p> <p>LCP 4.5/5.0 Narrow Plates LCP T-Plates 4.5/5.0 LCP 4.5/5.0 Broad Plates LCP 4.5/5.0 Broad Curved Plates LCP Reconstruction Plates 4.5/5.0</p> <p>In general the LCP Generic Plates & Instruments (Large Fragment) are indicated for the osteosynthesis of fractures at the Femur, Tibia, Humerus and Pelvic at proximal, distal and shaft areas.</p>	No contraindication specific to these devices.
LCP Low Bend Medial Distal Tibia Plates 3.5 mm	The Synthes LCP Low Bend Medial Distal Tibia Plates are intended for fixation of complex intra- and extra-articular fractures and osteotomies of the distal tibia, as a part of the Synthes Small Fragment LCP System.	No contraindication specific to these devices.
LCP Medial Proximal Tibial Plate 3.5	<p>The LCP Medial Proximal Tibial Plates 3.5 are indicated to buttress metaphyseal fractures of the medial tibial plateau, split-type fractures of the medial tibial plateau, medial split fractures with associated depressions and split or depression fractures of the medial tibial plateau.</p> <p>The plates may also be used for fixation of the proximal quarter (lateral and medial) of the tibia, as well as segmental fractures of the proximal tibia.</p>	No contraindication specific to these devices.

Systems	Indications	Contraindications
LCP Medial Proximal Tibial Plate 4.5/5.0	<p>The LCP Medial Proximal Tibial Plates 4.5/5.0 are indicated to buttress metaphyseal fractures of the medial tibial plateau, split-type fractures of the medial tibial plateau, medial split fractures with associated depressions and split or depression fractures of the medial tibial plateau. The plates may also be used for fixation of the proximal quarter (lateral and medial) of the tibia, as well as segmental fractures of the proximal tibia.</p> <p>The LCP Medial Proximal Tibial Plates 4.5/5.0 may also be used for fixation of nonunions and malunions of the medial proximal tibia and tibia shaft, as well as opening and closing wedge tibial osteotomies.</p>	No contraindication specific to these devices.
LCP Metaphyseal Plate for distal medial tibia	<p>The LCP Metaphyseal Plate for distal medial tibia is a preshaped plate that allows optimal treatment of juxta-articular fractures of the distal tibia extending into the shaft area. This plate takes the following characteristics of the distal tibia into account:</p> <ul style="list-style-type: none"> – Thin, soft tissue coverage – Complex anatomic shape of the bone 	No contraindication specific to these devices.
LCP Metaphyseal Plates	<p>The LCP Metaphyseal Plates are indicated to fix extra-articular fractures in the metaphyseal area that can extend into the shaft area. The 3.5 mm plates are indicated to fix fractures of the distal humerus and distal fibula. The 4.5/5.0 plates are indicated to fix fractures of the proximal humerus and distal tibial.</p>	No contraindication specific to these devices.
LCP Olecranon Plate	<ul style="list-style-type: none"> – Complex extra- and intra-articular olecranon fractures – Pseudoarthroses of the proximal ulna – Osteotomies – Simple olecranon fractures 	<ul style="list-style-type: none"> – Acute infections – Children in growth phase
LCP Pediatric Plates System	<p>The LCP Pediatric Plate System consists of different plates for different indications.</p> <p>The LCP Pediatric Hip Plate 2.7 is intended for use in infants up to three years, depending on body weight and bone quality.</p> <ul style="list-style-type: none"> – Neglected dislocation of the hip in combination with open reduction – Developmental coxa valga – Severe hip dysplasia <p>The LCP Pediatric Hip Plate for varus osteotomies is intended for use in pediatric patients up to adolescence and for small-stature adult patients.</p> <p>Specific indications include:</p> <ul style="list-style-type: none"> – Idiopathic valgus hip – Idiopathic and acquired subluxation of the femoral head – Femoral head subluxation in neuromuscular diseases/problems – High retroversion and anteversion in combination with a high CCD-angle <p>The LCP Pediatric Hip Plate for valgus osteotomies is intended for use in pediatric patients up to adolescence and for small stature adult patients.</p> <p>Specific indications include:</p> <ul style="list-style-type: none"> – High riding of greater trochanter and low shortening of the leg – Perthes' disease – Congenital pseudarthrosis of the femoral neck – Deformity of SCFE (Slipped Capital Femoral Epiphysis) – PFFD (Proximal Femoral Focal Deficiency) – Idiopathic coxa vara – Posttraumatic pseudarthrosis of the femoral neck <p>LCP Pediatric Hip Plate (3.5 & 5.0) 120° & 130° for fracture treatment and rotation correction is indicated for trans-trochanteric fractures with sufficient medial support, and femoral neck fractures Type I to III (see AO fracture classification).</p> <p>The LCP Pediatric Condylar Plate is intended for use in pediatric patients up to adolescence and for smallstature adult patients.</p> <p>Specific indications include:</p> <ul style="list-style-type: none"> – Fixed flexion contracture of knee in neurological conditions – Deformity correction in the distal femur – Rotational malalignment of the femur (if distal correction preferred) – Supracondylar fractures of the femur 	No contraindication specific to these devices.
LCP Periarticular Proximal Humerus Plate 3.5	<p>The LCP Periarticular Proximal Humerus Plate 3.5 addresses complex fractures of the proximal humerus.</p> <p>The 3.5 mm LCP Periarticular Proximal Humerus Plate is indicated for fractures and fracture dislocations (Neer type 2, 3, and 4-part fractures), osteotomies, and nonunions of the proximal humerus, particularly for patients suffering from osteoporosis.</p>	No contraindication specific to these devices.

Systems	Indications	Contraindications
LCP Posterior Medial Proximal Tibial Plate 3.5	The Synthes LCP Posterior Medial Proximal Tibial Plate 3.5 is indicated for internal fixation of posteromedial proximal tibia fractures including buttressing of fractures of the proximal, distal and metaphyseal areas of the tibia.	No contraindication specific to these devices.
LCP Proximal Femoral Hook Plate 4.5/5.0	The LCP Proximal Femoral Hook Plate 4.5/5.0 is intended for fractures of the femur including: <ul style="list-style-type: none"> – Fractures of the trochanteric region, trochanteric simple, cervico-trochanteric, trochanterodiaphyseal, multifragmentary pertrochanteric, intertrochanteric, reversed or transverse fractures of the trochanteric region or with additional fracture of the medial cortex – Fractures of the proximal end of the femur combined with ipsilateral shaft fractures – Metastatic fracture of the proximal femur – Osteotomies of the proximal femur – Also for use in fixation of osteopenic bone and fixation of non-unions or malunions – Periprosthetic Fractures 	No contraindication specific to these devices.
LCP Proximal Femoral Plate 4.5/5.0	The LCP Proximal Femoral Plate 4.5/5.0 is intended for fractures of the femur including: <ul style="list-style-type: none"> – Fractures of the trochanteric region, trochanteric simple, cervicotrochanteric, trochanterodiaphyseal, multifragmentary pertrochanteric, intertrochanteric, reversed or transverse fractures of the trochanteric region or with additional fracture of the medial cortex – Fractures of the proximal end of the femur combined with ipsilateral shaft fractures – Metastatic fracture of the proximal femur – Osteotomies of the proximal femur – Also for use in fixation of osteopenic bone and fixation of non-unions or malunions – Periprosthetic Fractures 	No contraindication specific to these devices.
LCP Proximal Radius Plates 2.4	Extra-articular and intra-articular fractures of the proximal radius and multifragmented radial neck fractures.	No contraindication specific to these devices.
LCP Proximal Tibial Plate 3.5	<ul style="list-style-type: none"> – Split-type fractures of the lateral tibial plateau – Lateral split fractures with associated depressions – Pure central depression fractures – Split or depression fractures of the medial plateau 	Isolated shaft fractures.
LCP Superior Anterior Clavicle Plate	<ul style="list-style-type: none"> – Fractures of the clavicle shaft – Fractures of the lateral clavicle – Malunions of the clavicle – Non-unions of the clavicle 	No contraindication specific to these devices.
LCP Superior Clavicle Plate	<ul style="list-style-type: none"> – Fractures of the clavicle shaft – Fractures of the lateral clavicle – Malunions of the clavicle – Non-unions of the clavicle 	No contraindication specific to these devices.
LCP Ulna Osteotomy System 2.7	Primary ulnar impaction syndrome <ul style="list-style-type: none"> – Degenerative triangular fibrocartilage complex (TFCC) tears – Lunotriquetral tears Secondary ulnar impaction syndrome <ul style="list-style-type: none"> – Incongruency (length discrepancy) of the distal radial-ulnar joint following distal radius fracture – Traumatic triangular fibrocartilage complex (TFCC) tears 	No contraindication specific to these devices.
LCP Volar Column Distal Radius Plates 2.4 mm	The LCP Volar Column Distal Radius Plate 2.4 mm is indicated for the fixation of complex intra- and extra-articular fractures and osteotomies of the distal radius.	No contraindication specific to these devices.
LCP Wrist Fusion Set	The LCP Wrist Fusion System is indicated for wrist arthrodesis and fractures of other small bones of the carpus. Specific indications include: <ul style="list-style-type: none"> – Posttraumatic arthrosis of the joints of the wrist – Rheumatoid wrist deformities requiring restoration – Complex carpal instability – Postseptic arthritis of the wrist – Severe unremitting wrist pain related to motion – Brachial plexus nerve palsies – Tumor resection – Spastic deformities 	No contraindication specific to these devices.

Systems	Indications	Contraindications
LISS and LCP DF	LCP DF is indicated for the stabilization of fractures of the distal femur. These include: <ul style="list-style-type: none"> – Distal shaft fractures – Supracondylar fractures – Intra-articular fractures – Periprosthetic fractures 	No contraindication specific to these devices.
LISS and LCP PLT	LCP PLT and LISS PLT are indicated for the stabilization of fractures of the proximal tibia. These include: <ul style="list-style-type: none"> – Proximal shaft fractures – Metaphyseal fractures – Intra-articular fractures – Periprosthetic fractures 	No contraindication specific to these devices.
Locking Attachment Plate	– Periprosthetic femoral shaft fractures: <ul style="list-style-type: none"> – Vancouver B – Vancouver C – Fractures around intramedullary implants 	No contraindication specific to these devices.
Midfoot Fusion Bolt Ø 6.5 mm	Treatment of gross collapse of the mid/hindfoot associated with: <ul style="list-style-type: none"> – Charcot foot – Neuropathic disease <p>Charcot foot can develop in several areas of the foot. The most common is the Lisfranc area, secondarily the Chopart area, thirdly the ankle area and fourthly the calcaneus – subtalar area. It often develops in the forefoot. The idea of the Midfoot Fusion Bolt is to treat the Lisfranc and Chopart area.</p>	<ul style="list-style-type: none"> – Peripheral vascular disease – Serious and life-threatening co-morbidities
Pelvic Implants and Instruments	<p>Symphysis fractures:</p> <ul style="list-style-type: none"> – Pubic Symphysis Plates <p>Pelvic brim fractures:</p> <ul style="list-style-type: none"> – Reconstruction Plates – Straight Plates – Curved Plates (88 mm radius, 108 mm radius) – J-Plates – Locking Reconstruction Plates 3.5 – Straight Plates – J-Plates – Reconstruction Plates 3.5 with wide Angle – Straight Plates <p>Ilium/Iliac wing fractures:</p> <ul style="list-style-type: none"> – Reconstruction Plates – Straight Plates – Curved Plates (88 mm radius, 108 mm radius) – Locking Reconstruction Plates 3.5 – Straight Plates – Low Profile Reconstruction Plates 3.5 with wide Angle – Straight Plates <p>Acetabulum fractures:</p> <ul style="list-style-type: none"> – Reconstruction Plates – Straight Plates – Curved Plates (88 mm radius, 108 mm radius) – Reconstruction Plates 3.5 with wide Angle – Straight Plates 	In-situ Plate Contouring Instruments are contraindicated for*: Plates with locking and coaxial combi-holes. In particular: <ul style="list-style-type: none"> – Locking Reconstruction Plates 3.5 – Pubic Symphysis Plates 3.5 – Spring Plates 3.5 – Reconstruction Plates 3.5 with coaxial combi-holes – Other plate sizes than 3.5 mm
Periarticular Aiming Arm Instruments for LCP Proximal Tibial Plate 4.5/5.0	The Synthes LCP Proximal Tibial Plates 4.5/5.0 are indicated for treatment of nonunions, malunions and fractures of the proximal tibia including: <ul style="list-style-type: none"> – Simple fractures – Comminuted fractures – Lateral wedge fractures – Depression fractures – Medial wedge fractures – Bicondylar, combination of lateral wedge and depression fractures – Fractures with associated shaft fractures 	No contraindication specific to these devices.

Systems	Indications	Contraindications
PHILOS and PHILOS Long	<p>PHILOS indications</p> <ul style="list-style-type: none"> – Dislocated two-, three-, and four-fragment fractures of the proximal humerus, including fractures involving osteopenic bone – Pseudarthroses in the proximal humerus – Osteotomies in the proximal humerus <p>PHILOS long indications</p> <ul style="list-style-type: none"> – As for PHILOS, but for fractures extending to the shaft or without medial support 	Not Available.
PHILOS with Augmentation	<p>PHILOS indications</p> <ul style="list-style-type: none"> – Dislocated two-, three-, and four-fragment fractures of the proximal humerus, including fractures involving osteopenic bone – Pseudarthroses in the proximal humerus – Osteotomies in the proximal humerus <p>PHILOS long indications</p> <ul style="list-style-type: none"> – As for PHILOS, but for fractures extending to the shaft or without medial support <p>PHILOS Augmentation indications</p> <ul style="list-style-type: none"> – As for PHILOS and PHILOS long, but exclusively in conjunction with osteoporotic bone – The perforated screws may also be used without cement augmentation 	PHILOS Augmentation contraindications
Quadrilateral Surface Plates 3.5	The Synthes Quadrilateral Surface Plates 3.5 are indicated for quadrilateral surface comminution associated with acetabular fractures when used in conjunction with Synthes Pelvic Reconstruction Plates.	No contraindication specific to these devices.
Rotation Correction Plates 1.5 and 2.0	<p>1. All fractures of the phalanges and metacarpals, where the exact reposition is difficult or where a rotational error can easily occur.</p> <ul style="list-style-type: none"> – subcapital fractures of the metacarpals (in particular impacted compression fractures) – transverse fractures – short oblique fractures – comminuted fractures – defect fractures (circular saw injuries) – amputation injuries to the fingers (with primary shortening) – Winterstein fracture, Rolando fracture <p>2. Corrective osteotomies for fractures of the phalanges or metacarpals that have healed with axial errors and/or rotational errors. In the case of metaphyseal fractures, comminuted fractures and osteoporotic bone, the clinical results can be improved by the angular-stable screw/plate connection.</p>	No contraindication specific to these devices.
Sacral Bars	The Synthes Sacral Bars are indicated for fixation of the posterior pelvis, in areas of the posterior superior iliac spine, posterior inferior iliac spine, for sacral fractures, and fracture dislocations of the sacroiliac joint.	No contraindication specific to these devices.
Slipped Capital Femoral Epiphysis (SCFE) Screw System	<p>The Slipped Capital Femoral Epiphysis (SCFE) Screws are indicated for:</p> <ul style="list-style-type: none"> – Femoral neck fractures in children, adolescents and adults – Slipped capital femoral epiphysis – Tibial plateau fractures – Ankle arthrodesis – Intercondylar fractures – Sacroiliac joint disruptions – Subtalar arthrodesis <p>This device is not approved for screw attachment or fixation to the posterior elements (pedicles) of the cervical, thoracic, or lumbar spine.</p>	No contraindication specific to these devices.
Spring Plates 3.5	The Synthes 3.5 mm Spring Plate is intended for pelvic and acetabular reconstructive surgery	No contraindication specific to these devices.
Standard DHS Lag Screw with LCP DHHS Sideplate	<ul style="list-style-type: none"> – Pertrochanteric fractures of type 31-A1 and 31-A2 – Intertrochanteric fractures of type 31-A3 – Basilar neck fractures 31-B (DHS Screw in conjunction with an antirotation screw) – Subtrochanteric fractures 	The DHS is not to be used in cases where there is a high incidence of: <ul style="list-style-type: none"> – Sepsis – Malignant primary or metastatic tumors – Material sensitivity – Compromised vascularity

Systems	Indications	Contraindications
Surgical Technique – VA LCP Two-Column Volar Distal Radius Plate 2.4/2.7mm, Extra-long	The DePuy Synthes VA-LCP Two-Column Volar Distal Radius Plate 2.4/2.7 mm, Extra-Long is intended for intra- and extra-articular fractures, osteotomies, nonunions and malunions of the distal radius, with or without extension into the radial diaphysis.	No contraindication specific to these devices.
The Calcaneal Plate	The calcaneal plates address fractures of the calcaneus. The calcaneal plate is indicated for fractures and osteotomies of the calcaneus including, but not limited to, extra-articular, intra-articular, joint depression and tongue type fractures.	No contraindication specific to these devices.
The Locking Calcaneal Plate	The locking calcaneal plates address complex fractures of the calcaneus. The locking calcaneal plate is indicated for fractures and osteotomies of the calcaneus including, but not limited to, extra-articular, intra-articular, joint depression, tongue type, and severely comminuted fractures.	No contraindication specific to these devices.
Tomofix	<p>Tomofix Medial High Tibia and Medial High Tibia Small Stature Plate: Open-wedge and closed-wedge osteotomy of the medial proximal tibia for the treatment of:</p> <ul style="list-style-type: none"> – Unicompartmental medial or lateral gonarthrosis with malalignment of the proximal tibia – Idiopathic or posttraumatic varus or valgus deformity of the proximal tibia <p>Tomofix Lateral High Tibia Plate: Open-wedge and closed-wedge osteotomy of the lateral proximal tibia for the treatment of:</p> <ul style="list-style-type: none"> – Unicompartmental medial or lateral gonarthrosis with malalignment of the proximal tibia – Idiopathic or posttraumatic varus or valgus deformity of the proximal tibia <p>Tomofix Lateral Distal Femur Plate: Open-wedge and closed-wedge osteotomy of the lateral distal femur for the treatment of:</p> <ul style="list-style-type: none"> – Unicompartmental medial or lateral gonarthrosis with malalignment of the distal femur – Idiopathic or posttraumatic varus or valgus deformity of the distal femur 	Inflammatory arthritis
Tomofix Medial Distal Femur (MDF)	Closed-wedge osteotomies of the medial distal femur for the treatment of:	Inflammatory arthritis
Tomofix Medial High Tibial Plate (MHT)	Open-wedge and closed-wedge osteotomies of the medial proximal tibia for the treatment of:	Inflammatory arthritis
VA-LCP Ankle Trauma System 2.7/3.5	<p>VA-LCP Medial Distal Tibial Plate 2.7/3.5 Indicated for complex intra- and extra-articular fractures of the distal tibia.</p> <p>VA-LCP Anteromedial Distal Tibial Plate 2.7/3.5 Indicated for complex intra- and extra-articular fractures of the distal tibia.</p> <p>VA-LCP Anterolateral Distal Tibial Plate 2.7/3.5 Indicated for complex intra- and extra-articular fractures of the distal tibia.</p> <p>VA-LCP Lateral Distal Fibula Plate 2.7 Indicated for fractures and non-unions of the metaphyseal and diaphyseal region of the distal fibula, especially in osteopenic bone.</p> <p>VA-LCP Posterolateral L- and T-Plates 2.7 Indicated for buttressing of partial articular fractures and bone fragments of the distal tibia.</p>	No contraindication specific to these devices.

Systems	Indications	Contraindications
VA-LCP Anterior Clavicle Plate	<ul style="list-style-type: none"> – Fractures of the clavicle shaft – Fractures of the lateral clavicle – Malunions of the clavicle – Non-unions of the clavicle 	No contraindication specific to these devices.
VA-LCP Condylar Plate 4.5/5.0	The Synthes VA-LCP Condylar Plate 4.5/5.0 system is indicated for buttressing multifragmentary distal femur fractures including: supra-condylar, intra-articular and extra-articular condylar fractures, periprosthetic fractures, fractures in normal or osteopenic bone, nonunions and malunions.	No contraindication specific to these devices.
VA-LCP Distal Humerus Plates 2.7/3.5	<ul style="list-style-type: none"> – Intra-articular fractures of the distal humerus – Supracondylar fractures of the distal humerus – Nonunions of the distal humerus – Osteotomies of the distal humerus(e.g. due to malunions, deformities) 	No contraindication specific to these devices.
VA LCP® Medial Column Fusion Plates 3.5	The DePuy Synthes VA LCP Medial Column Fusion Plates 3.5 are indicated for deformities, severe arthritis, and arthrosis of the medial column consisting of the first metatarsal, medial cuneiform, navicular and talus.	No contraindication specific to these devices.
VA-LCP Olecranon Plates 2.7/3.5	<p>VA-LCP Proximal Olecranon Plates</p> <ul style="list-style-type: none"> – Fractures of the proximal olecranon – Osteotomies of the olecranon for distal humerus fracture treatment <p>VA-LCP Olecranon Plates</p> <ul style="list-style-type: none"> – Intra-articular fractures of the olecranon including fractures extending into the coronoid – Nonunions of the olecranon – Osteotomies of the olecranon (e.g. due to malunions, deformities) <p>VA-LCP Proximal Ulna Plate, extra-articular</p> <ul style="list-style-type: none"> – Extra-articular fractures of the proximal ulna – Nonunions of the proximal ulna – Osteotomies of the proximal ulna (e.g. due to malunions, deformities) 	No contraindication specific to these devices.
VA-LCP Proximal Tibial Plate 3.5	<p>Fractures of the proximal tibia in adults and adolescents with closed growth plates including</p> <ul style="list-style-type: none"> – Proximal split, depression or split-depression fractures – Bicondylar or pure metaphyseal fractures – Associated metaphyseal or associated shaft fractures – Periprosthetic fractures 	No contraindication specific to these devices.
VA Locking Calcaneal Plates 2.7	The Synthes Variable Angle Locking Calcaneal plates 2.7 are indicated for intra- and extra-articular fractures of the calcaneus, as well as deformities and malunions	No contraindication specific to these devices.
VA-Locking Intercarpal Fusion System	<p>Indicated for fusion of small bones of the hand including:hamate, capitate, lunate, and triquetrum, for the revision offailed partial wrist fusions, and is indicated for use in patients suffering pain and/or loss of function due to:</p> <ul style="list-style-type: none"> – Osteoarthritis – Rheumatoid arthritis – Post-traumatic or degenerative wrist arthritis – Carpal instability 	No contraindication specific to these devices.
Variable Angle LCP 1st MTP Fusion Plates 2.4/2.7	The 1 st MTP Fusion Plate of the Variable Angle LCP Forefoot/Midfoot System 2.4/2.7 is indicated for deformations of the first metatarsophalangeal (MTP) joint (Hallux Rigidus) and fractures, nonunions and replantations of the first metatarsal bone, particularly in osteopenic bone.	No contraindication specific to these devices.
Variable Angle LCP Dorsal Distal Radius Plate 2.4	<p>2.4 mm Variable Angle LCP Dorsal Distal Radius Plates are indicated for:</p> <ul style="list-style-type: none"> – Dorsally displaced fractures – Extra-articular and intra-articular fractures with metaphyseal defect – Open joint reconstruction – Combination of distal radius with carpal and metacarpal fractures – Corrective osteotomies after distal radius malunion 	No contraindication specific to these devices.
Variable Angle LCP Forefoot/Midfoot System 2.4/2.7	The Straight Fusion Plates, T-Fusion Plates, L-Fusion Plates, Cloverleaf Fusion Plates and X-Plates of the Variable Angle LCP Forefoot/Midfoot System 2.4/2.7 are indicated for fractures, deformations, revisions and replantations of bones (e.g. tarsals, metatarsals and phalanges) and bone fragments, particularly in osteopenic bone.	No contraindication specific to these devices.

Systems	Indications	Contraindications
Variable Angle LCP Mesh Plate 2.4/2.7	The Mesh Plate, part of the Variable Angle LCP Forefoot/Midfoot System 2.4/2.7, is indicated for fractures, deformations, severe osteoarthritis and non- and mal-unions in the forefoot and midfoot, particularly in osteopenic bone.	No contraindication specific to these devices.
Variable Angle LCP Opening Wedge Plates 2.4/2.7	The Opening Wedge Plate of the Variable Angle LCP Forefoot/Midfoot System 2.4/2.7 is indicated for deformations (e.g. hallux valgus), nonunions, and replantations of the first metatarsal bone, particularly in osteopenic bone.	No contraindication specific to these devices.
Variable Angle LCP Tarsal Plates 2.4/2.7	The Synthes VA locking tarsal plates are intended for the fixation of fractures, osteotomies, nonunions, replantations, and fusions of the Cuboid and Navicular bones, particularly in osteopenic bone.	No contraindication specific to these devices.
Variable Angle LCP TMT Fusion Plates 2.4/2.7	The 1 st TMT Fusion Plates of the Variable Angle LCP Forefoot/Midfoot System 2.4/2.7 are indicated for deformations of the first tarsometatarsal (TMT) joint (Hallux valgus) and fractures, nonunions and replantations of the first metatarsal bone, particularly in osteopenic bone. The TMT Fusion Plate of the Variable Angle LCP Forefoot/Midfoot System 2.4/2.7 is indicated for deformations of the second and third tarsometatarsal (TMT) joint and fractures, deformations, nonunions and replantations of the second and third metatarsal bone, particularly in osteopenic bone.	No contraindication specific to these devices.
Variable Angle LCP Two-Column Volar Distal Radius Plate 2.4	Variable Angle LCP Two-Column Volar Distal Radius Plates 2.4 are indicated for the fixation of intra- and extra-articular fractures and osteotomies of the distal radius.	No contraindication specific to these devices.
Variable Angle LCP Volar Rim Distal Radius Plate 2.4	Variable Angle LCP Volar Rim Distal Radius Plate 2.4 is indicated for the fixation of complex intra-articular and extra-articular fractures of the distal radius.	No contraindication specific to these devices.
Variable Angle Locking Hand System	The Variable Angle Locking Hand System is indicated for the treatment of fractures, deformities and degenerative diseases in the hand.	No contraindication specific to these devices.

CE
0123



Synthes GmbH
Eimattstrasse 3
4436 Oberdorf
Switzerland
Tel: +41 61 965 61 11
Fax: +41 61 965 66 00
www.depuysynthes.com